

***EU DECLARATION OF
CONFORMITY FOR
ATTACHMENT OF 'CE'
MARK***



***DECLARATION OF
CONFORMITY FOR
ATTACHMENT OF
'UKCA' MARK***



We, the manufacturer: Ecolab Limited.
52 Royce Close,
West Portway
Andover, UK SP10 3TS
Tel: +44 (0) 1264 - 835835

declare under our sole responsibility that the appliance:

Bioquell R-30

Serial Number as stated on the unit's rating plate

complies with the Essential Requirements of the Machinery Directive 2006/42/EC, EMC Directive 2014/30/EU, the Low Voltage Directive 2014/35/EU and Restriction of Hazardous Substances in Electrical and Electronic Equipment 2011/65/EU by means of the following harmonised standards:

complies with the statutory requirements of the Supply of Machinery (Safety) Regulations 2008, EMC Regulations 2016, Electrical Equipment (Safety) Regulations 2016, The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 by means of the following designated standards:

Machinery Directive / Regulation: by means of BS EN 12100:2010

EMC Directive / Regulation: by means of BS EN 61326-1:2013

Low Voltage Directive / Electrical Equipment Regulation: by means of BS EN 61010-1:2010+A1:2019

Restriction of Hazardous Substances in Electrical and Electronic Equipment by means of BS EN IEC 63000:2018.

A copy of the Technical File for this equipment is available from: Quality Manager

Ecolab SNC, 153 Quai du Rancy, 94380
Bonneuil-sur-Marne, Paris, FRANCE

Ecolab Limited, 52 Royce Close, West
Portway, Andover, SP10 3TS, UK

Authorised signatory of manufacturer:

O.R.Cumberlege, Bioquell R&D Group Leader

27th January 2022